



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska


JUN 03 2009

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program


Supervisor/Operator Performing the Verification Procedure:		DataMaster cdm S/N <u>130119</u> ✓
Name <u>Tobin P Brennan</u>		ID# <u>4542</u> Date <u>6-1-09</u> ✓
A	Agency <u>Soldotna Police Department</u>	Phone # <u>(907) 262-4455</u> ✓
Instrument Location <u>Soldotna Police Department (44510 Sterling Hwy, Soldotna Alaska 99669)</u> ✓		
<hr/>		
B	Alco S/N <u>X301573</u> ✓	Target Value <u>.080</u> ✓ High Pressure <u>1400</u>
Alco Test Values		
	<u>.080</u> ✓ 1 st Alco	<u>.081</u> ✓ 2 nd Alco
Signature 		 6/1/09 6/5/09
(OVER)		


I, Nita J. Bolz, after being first duly sworn, depose and state as follows: (Do Not write in the area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.


Nita J. Bolz
Scientific Director
State Breath Alcohol Program

Subscribed and sworn before me this 12 day of June, 2009.


Carolyn M. Noland
Notary Public, State of Alaska
Commission Expires with Office



VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130119

(CONTINUED FROM FRONT PAGE)

C TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130119

JUNE 01, 2009

OPERATOR'S NAME: *cap 6/6/09*
BRENNAN
OPERATOR'S NUMBER: 4542
SUBJECT'S LAST NAME:
VERIFICATION OF CAL
SUBJECT'S FIRST NAME/MI :
UOC

O.L. #: UOC
DEPT/AGENCY: SLD1
CASE/REPORT: UOC
TEST TYPE: U
ALCO TARGET VALUE: .080 ✓
ALCO S/N: K301573 ✓

--- BREATH ANALYSIS ---

.080 ADJUSTED FOR 30.32 in ✓
ALCO TARGET .081 10:33
BLANK TEST .000 10:34
INTERNAL STANDARD VERIFIED 10:34
ALCO TO 30.32 in .080 ✓ 10:34
BLANK TEST .000 10:35
SUBJECT SAMPLE *cap 6/6/09* .080 10:36
BLANK TEST .000 10:36
ALCO TO 30.32 in .081 ✓ 10:37
BLANK TEST .000 10:38

DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY
DATAMASTER cdm 130119

JUNE 01, 2009 ✓
TIME 10:28

--- DIAGNOSTIC CHECK ---

COMPUTER: OKAY
PROGRAM: OKAY
SOFTWARE DATE: 03/29/00
HEATERS
SAMPLE CHAMBER: 43c
BREATH TUBE: 45c
BAROMETER: 30.32 in
FLOW DETECTOR: OKAY
PUMP
HIGH SPEED: OKAY
DETECTOR: OKAY
FILTERS: OKAY
QUARTZ STANDARD: OKAY
CALIBRATION: OKAY

PRINTER TEST

!@#%&'()*+,-./:0123456789:K=>?@AB0DEFG
HIJKLMNOPQRSTUVWXYZ0123456789:K=>?@AB0DEFG
QRSTUVWXYZ0123456789:K=>?@AB0DEFG